


PRODUCT DATA SHEET

Revision 1	PCR 12-Strip Tube, With Cap	
Validfor Item-No.:	2050010/2050110/2051010/2051110	

1.	Description / Specification	
1.1	Description	PCR 12-Strip Tube, With Cap 2050010 PCR 12-strip Tube, 0.1 ml, clear with lid 2050110 PCR 12-strip Tube, 0.1 ml, white with lid 2051010 PCR 12-strip Tube, 0.2 ml, clear with lid 2051110 PCR 12-strip Tube, 0.2 ml, white with lid
1.2	Dimensions	Tube: see Customer Drawing
1.3	Volume	Max. volume: 0.1 ml / 0.2 ml Working volume: 0.1 ml / 0.2 ml
1.4	Material / Resin	<u>Tube:</u> PP (Polypropylene) <u>Cap:</u> PP (Polypropylene) The materials for manufacturing are USP Class VI
1.5	Colour	<u>Tube:</u> clear or white <u>Cap:</u> clear
1.6	Sterilization	No
1.7	Quality Control	- Raw Material-Control: physical testing - Product-Control: testing of attributive and variable characteristics in accordance with the valid specification
1.8	Intended Use	General laboratory product for molecular biology to be used by qualified personnel in a laboratory environment.
1.9	Other Information	- For single use only - Expiry date and Lot-No. printed on the Packaging

2.	Features	
2.1	Basic features	Free of detectable DNase/RNase, human DNA and pyrogens. Contents non-cytotoxic
2.2	Temperature range	For application: -20°C to +105°C
2.3	Autoclavability	Not recommended
2.4	Centrifugation, max. RCF	N/A
2.5	Shelf life	3 years
2.6	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	125
3.2	Pieces / Box	1250
3.3	Lot-No.	DDD YY XXX (day, year, product No.)
3.4	Other Information	Certificate of Quality to download

4.	Other Information	
4.1	Laboratory use only	

Prior Issue	Drawn	Approved	Released	CONFIDENTIAL: Information contained in this document or drawing is confidential and proprietary to Zhejiang Saining Biotechnology Co., Ltd. This document may not be reproduced for any reason without written permission from Zhejiang Saining Biotechnology Co., Ltd. All rights of design, invention, and copyright are reserved.
Revision 1	Date 07.05.2024	Date 14.05.2024	Date 31.05.2024	
Date 01.06.2024	Name Chen Anjing	Name Zhang Heng	Name Wang Kedong	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.